

West Side T6 2



WASHINGTON STATE PATROL
SCALE TEST REPORT AND CERTIFICATION OF INSPECTION

DATE: July 10, 2006 INITIALS: DEC OWNER: WSP
DISTRICT: 2 ADDRESS: I-5 South MP 141
LOCATION: Seaside South T6 27 COUNTY: King
MAKE: Volvo TYPE: truck scale ACCURACY CLASS: III
MODEL: 8530 ID. NO: 4338197-4NV CAPACITY: 100000 SECTIONS: 2
MINIMUM GRADUATIONS: 20 REG. ELEMENT: Digital PLATFORM: 12' X 20'

LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR	LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR
<u>Sec 1</u>	<u>20000</u>	<u>20,000</u>	<u>0</u>				
<u>Center</u>	<u>20000</u>	<u>20000</u>	<u>0</u>				
<u>Sec 2</u>	<u>20000</u>	<u>20020</u>	<u>+20</u>				

Using testing procedures set forth in Handbook 44 promulgated by the National Institute of Standards and Technology and test weights which were certified as accurate as is shown on the attached "Certificate of Accuracy", No. 2 2441 - Load 2

I do hereby certify under penalty of perjury under the laws of the state of Washington that the above described weighing device met or exceeded the standards of accuracy.
Signed by: David E. Brown Date: 8/3/2006



State of Washington
County of THURSTON
Signed and sworn to (or affirmed) before me on 8-3-2006 by
Sylvia S. Lewallen
(Signature of notary public)
SYLVIA S. LEWALLEN
Printed Name of Notary
My appointment expires 3-28-2007